

## Calibration Services for Digital Barometers PTB210

Please fill in the model, serial number and order code with the selected options:

<b>Model:</b>	<b>Serial no:</b>	<b>Code: 223051</b>		<b>A</b>	<b>A</b>	<b>A</b>
<b>Calibration class</b>	ISO/IEC 17025 Accredited calibration ISO 10012 Compliant calibration	<b>A</b> <b>B</b>				
<b>Pressure calibration</b> <small>NOTE 1</small>	50, 75, 150, 300, 500, 700, 900, 1000 and 1100 hPa 500, 550, 650, 750, 850, 950, 1000, 1050 and 1100 hPa 50, 100, 200, 400, 600, 800, 1000, 1200 and 1300 hPa 300, 450, 600, 750, 900, 1000, 1100, 1150 and 1200 hPa 800, 850, 900, 950, 1000, 1050 and 1100 hPa 300, 450, 600, 750, 900, 1000, 1100, 1150, 1200 and 1300 hPa	<b>A</b> <b>B</b> <b>C</b> <b>D</b> <b>E</b> <b>F</b>				
<b>Calibration temperature</b>	+23 °C (room temperature)	<b>A</b>				
<b>Adjustment</b>	No adjustment Adjusted at 1000 hPa pressure (offset) Adjusted using 8 point MPC correction Adjusted using 11 point MPC correction	<b>A</b> <b>B</b> <b>C</b> <b>D</b>				
<b>Calibration method</b>	Measurement via digital interface	<b>A</b>				
<b>Results reported, pressure</b>	hPa kPa hPa and inHg hPa and mmHg hPa and torr hPa and mmH <sub>2</sub> O hPa and psia	<b>A</b> <b>B</b> <b>C</b> <b>D</b> <b>E</b> <b>F</b> <b>G</b>				
<b>Language of the Certificate</b>	Certificate in English	<b>A</b>				

**NOTE 1:** Selected calibration range shall not exceed the measurement range of the transmitter.

Please fill in the end customer information for Certificate:

End customer	
Order received and approved	