VAISALA

Print this order form
Fill out the required information
Pack instrument carefully
Include filled form into the package
Send to Vaisala

P.O. Box 26, FI-00421 Helsinki, Finland
Tel. +358 9 894 91
Fax. +358 9 8949 2790
Email: helpdesk@vaisala.com
www.vaisala.com

Calibration Services for PTB330TS Transfer Standard

Please fill in the serial number(s) and order code with the selected options:

Model (transmitter): PTB330TS	**Serial no:**													
Model (probe, if any): HMP155	**Serial no:**	**Code: 223943**			**A**				**A**			**A**		
Calibration class		ISO/IEC 17025 Accredited calibration	**A**											
		ISO 10012 Compliant calibration	**B**											
Pressure calibration		No calibration		**A**										
		500, 550, 650, 750, 850, 950, 1000, 1050 and 1100 hPa		**B**										
Calibration temperature		+23 °C (room temperature)			**A**									
Temperature calibration		No calibration/one-point calibration at +23 °C				**A**								
		0, +20 and +40 °C				**B**								
		+17.5, +20.0, +22.5, +25.0 and +27.5 °C				**C**								
		-20, 0, +20, +40 and +60 °C				**D**								
		-40, -20, 0, +20, +40 and +60 °C				**E**								
		0, +10, +20, +30, +40, +50 and +60 °C				**F**								
		-80, -60, -40, -20, 0, +20, +40 and +60 °C				**G**								
		-60, -40, -20, 0, +20, +40 and +60 °C				**H**								
		-40, -30, -20, -10, 0, +10, +20, +30 and +40 °C				**I**								
NOTE 1		-40, -20, 0, +20 and +40 °C				**J**								
Free point 1:	Free point 2:	Free point 3:				**K**								
Humidity calibration		No calibration					**A**							
NOTE 2		0.1, 11, 33, 53 and 75 %RH					**B**							
		0.1, 11, 33, 53, 75, 85 and 97 %RH					**C**							
Adjustment		No adjustment						**A**						
		Adjusted for optimum performance						**B**						
Calibration methods		Measurement via digital interface							**A**					
Results reported, pressure		hPa								**A**				
		kPa								**B**				
		hPa and inHg								**C**				
		hPa and mmHg								**D**				
		hPa and torr								**E**				
		hPa and mmH_2O								**F**				
		hPa and psia								**G**				
Results reported, temperature and humidity		%RH and °C									**A**			
		%RH, °C and °F									**B**			
Language of the Certificate		Certificate in English											**A**	

NOTE 1: Please fill in the selected free calibration points in range from -80 to +60 °C.
NOTE 2: Please note that the humidity calibration is performed using saturated salt solutions at room temperature.
Please note that one-point temperature calibration at +23 °C temperature is included into the humidity calibration.

Please fill in the end customer information for Certificate:

End customer	
Order received and approved	

VAISALA

Vaisala Oyj
P.O.Box 26
FI-00421 Helsinki
Finland
Tel. (+358 9) 894 91
Fax. (+358 9) 8949 2227

For other Vaisala
locations visit us at:
www.vaisala.com

Ref. DOC217194-D © 2009 Vaisala. All specifications are subject to change without notice